

**CHAPTER 33-11-05**  
**BASIC LIFE SUPPORT AIR AMBULANCE LICENSE**

Section

33-11-05-01

Training Standards for Primary Care Provider

33-11-05-02

Minimum Equipment Requirements

**33-11-05-01. Training standards for primary care provider.** The primary care provider must have current emergency medical technician license or its equivalent and must have current cardiopulmonary resuscitation certification.

**History:** Effective August 1, 2003; amended effective January 1, 2006.

**General Authority:** NDCC 23-27-04

**Law Implemented:** NDCC 23-27-04

**33-11-05-02. Minimum equipment requirements.**

1. Patient litter or stretcher.
2. One stethoscope.
3. One blood pressure cuff with aneroid gauge.
4. Manual suction device with catheter.
5. One set of oropharyngeal airways including six sizes, from infant through adult.
6. One set of nasopharyngeal airways.
7. Oxygen administration system, including a protective pressure gauge, a nongravity dependent flowmeter, supply tubing, a nonrebreather mask, and a nasal cannula. The unit must be capable of achieving an oxygen delivery flow rate of at least fifteen liters per minute for one hour.
8. Mouth-to-mask artificial ventilation device, with a supplemental oxygen inlet port such as a pocket mask, suitable for use on infant through adult patients. This may be replaced with bag valve mask devices with masks for infant, child, and adult patients.

**History:** Effective August 1, 2003.

**General Authority:** NDCC 23-27-04

**Law Implemented:** NDCC 23-27-04